



ČESKÝ METROLOGICKÝ INSTITUT

EC type-approval certificate

Number: TCM 128/06 - 4428

Revision 1

Issued by: Český metrologický institut
Okružní 31
638 00 Brno
Česká republika

Notified body: 1383

In accordance with: Annex II, article 1. of the Government Order No. 326/2002 Coll. and the Council Directive 90/384/EEC on non-automatic weighing instruments.

Issued to: AXIS Sp. z o.o.
80-125 Gdańsk
ul. Kartuska 375 B
Poland

In respect of: non-automatic weighing instrument, single interval
class II
Type: AG series
Max = 50g, 100 g, 200 g, 300 g, 500 g, 600 g, 1000 g, 2000 g, 3000 g, 4000 g
(depending on model)
 $e = 0,01 \text{ g} / 0,1 \text{ g}$
 $e = 10d$
 $n \leq 50000$
Temperature range + 18°C / + 33°C


Manufacturer: AXIS Sp. z o.o., Poland


Valid until: 11. September 2016

Document number: 0115-CS-A002-06

Description: Essential characteristics, approved conditions and special conditions, if any, are described in this EC type-approval certificate. This certificate contains 6 pages.

Brno, 5. January 2007




RNDr. Pavel Klenovský
Notified Body No. 1383